



SEA-BIRD ELECTRONICS, INC.

13431 NE 20th St. Bellevue, Washington 98005 USA

Phone: (425) 643-9866 Fax: (425) 643-9954 www.seabird.com

Pressure Test Certificate

Customer Atlantic Marine Center

Job Number 84233

Date 4/14/2015

Technician PS

Serial Number 09P47490-0905

Low Pressure (PSI) 50 PSI

Time (Minutes) 15 Minutes

High Pressure (PSI) 10000 PSI

Time (Minutes) 45 Minutes

Pass ☒

Fail ☐

Comments

Replaced the main piston "O"-Rings.

